



February 13, 2025

BLAISE DERRICO  
TREND MEDICAL LLC  
6583 RUCH RD  
BETHLEHEM, PA 18017

**Document Control Number (DCN): 24344003000003**

| Manufacturer Name | Product Name       | Model Number | Assigned HCPCS Code(s) |
|-------------------|--------------------|--------------|------------------------|
| TREND MEDICAL LLC | TREND PRO 71 ANKLE | DCT-1971-X   | L1971                  |
| TREND MEDICAL LLC | TREND PRO 71 ANKLE | DCT-1971-X   | L1906                  |

Dear BLAISE DERRICO,

The Pricing, Data Analysis, and Coding (PDAC) Contractor has reviewed the product(s) listed above and has approved the listed Healthcare Common Procedure Coding System (HCPCS) code(s) for billing the four Durable Medical Equipment Medicare Administrative Contractors (DME MACs).

The PDAC Contractor provides coding assistance to manufacturers to ensure proper coding of Durable Medical Equipment, Prosthetics, Orthotics, and Supplies (DMEPOS). The PDAC publishes coding decisions based on the coding guidelines established by the Local Coverage Determinations (LCDs) and associated Policy Articles and any related Advisory Articles established by the DME MACs. All products submitted to the PDAC for a coding verification review are examined by coders and professionals following a formal, standardized process.

Based on this review and application of DME MAC policy, the HCPCS code(s) listed below should be used when billing the DME MACs:

L1906 ANKLE FOOT ORTHOSIS, MULTILIGAMENTOUS ANKLE SUPPORT,  
PREFABRICATED, OFF-THE-SHELF

L1971 ANKLE FOOT ORTHOSIS, PLASTIC OR OTHER MATERIAL WITH ANKLE  
JOINT, PREFABRICATED, INCLUDES FITTING AND ADJUSTMENT

If you disagree with this decision, you may submit a new application form with evidence to substantiate the reconsideration of PDAC's coding determination. The [Application for Code Verification Request form](#) is located on the PDAC website at [www.dmepdac.com](http://www.dmepdac.com).

It is the responsibility of manufacturers and distributors to notify the PDAC immediately of any changes involving their products, as listed on the Product Classification List (PCL) on the Durable Medical Equipment Coding System (DMECS). Further information for requesting updates to the PCL can be found on the PDAC website at [www.dmepdac.com](http://www.dmepdac.com). It is also the responsibility of manufacturers and distributors to assure their websites and product marketing materials accurately reflect the product reviewed by the PDAC and the coding decision assigned.

An assignment of the HCPCS code(s) to product(s) is not an approval or endorsement of the product(s) by Medicare or Palmetto GBA; nor does it imply, or guarantee claim reimbursement or coverage.

If you have questions, please contact the PDAC HCPCS Helpline at (877) 735-1326 during the hours of 9:30 a.m. to 5:00 p.m. ET, Monday through Friday. You may also visit our [website](#) to chat with one of our representatives or select the Contact Us button at the top of the page for email, FAX, or postal mail information.

Sincerely,

Pricing, Data Analysis, and Coding (PDAC)  
Palmetto GBA, LLC  
[www.dmepdac.com](http://www.dmepdac.com)